

Test Report

Report No.: CTT2101031375EN

Date: Jan. 21, 2021

Page 1 of 2

Applicant: WE PLASTIC MFG CORPORATION LIMITED

Address: LIUGUANZHUANG,JUXIAN COUNTY,RIZHAO CITY,SHANGDONG PROVINCE,CHINA

Sample Received Date: Jan. 18, 2021

Completed Date: Jan. 21, 2021

The following merchandise was (were) submitted and identified on behalf of the applicant as:

Sample Name: TPE GLOVES

Test Result(s): Please refer to next page(s).

Test Requested and Conclusion(s):

| No. | Test Sample | Standard and Requirement | Conclusion(s) |
|-----|---------------------------------------|---|---------------|
| 1 | Tested materials of submitted samples | US FDA 21 CFR 177.2600: Rubber articles intended for repeated use. - Total extractives in Distilled water - Total extractives in n-Hexane | PASS |

Signed for and on Behalf of CTT



Frank / Lab manager

Consumer Testing Technology Co., Ltd.

This test report is issued by the company subject to its General Conditions of Services available on request and accessible at <http://www.cttlab.com/order/201905050400480820.pdf>. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues defined therein. Unless otherwise stated the results shown in this report refer only to the sample(s) received. Without prior written permission of the company, this test report cannot be reproduced, except in full. Any inquiry about this report, please raise from the date of receipt of the report within 30 days, overdue will not be accept. Items marked with "n" means they are not accredited by CNAS (if with CNAS logo), "s" means the item of subcontractor. All or Part of items of this report are not in the accredited scope of CMA and cannot be as domestic social impartiality proof data.

Consumer Testing Technology Co.,Ltd.

1st-2nd Floor, Building 5, District B, Yingcai Entrepreneurship Park, No.E21, Xinke Road, Yiwu, Zhejiang, China
Tel:86-579-8998 6543 Fax:86-579-8998 6547 Email: zj@cttlab.com Hotline:400 6789 666 [Http://www.cttlab.com](http://www.cttlab.com)



Test Report

Report No.: CTT2101031375EN

Date: Jan. 21, 2021

Page 2 of 2

Test Result(s):

US FDA 21 CFR 177.2600 - Total extractives

Method: With reference to US FDA 21 CFR 177.2600

| Material No. | Test Condition | Limit (mg/inch ²) | Result (mg/inch ²) | Conclusion |
|--------------|--|-------------------------------|--------------------------------|------------|
| 1 | Distilled water, reflux temperature for first 7 hours | 20 | 0.37 | PASS |
| | Distilled water, reflux temperature for succeeding 2 hours | 1 | 0.38 | PASS |
| | n-Hexane, reflux temperature for first 7 hours | 175 | 2.37 | PASS |
| | n-Hexane, reflux temperature for succeeding 2 hours | 4 | 0.37 | PASS |

- Note:**
1. mg/inch² = milligram per square inch.
 2. The test result(s) refers to CTT2101031374EN.

Test Material List

| Material No. | Description | Location |
|--------------|---------------------------|----------|
| 1 | Blue translucence plastic | Gloves |

Photo of Sample:



End of Report

This test report is issued by the company subject to its General Conditions of Services available on request and accessible at <http://www.cttlab.com/order/201905050400480820.pdf>. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues defined therein. Unless otherwise stated the results shown in this report refer only to the sample(s) received. Without prior written permission of the company, this test report cannot be reproduced, except in full. Any inquiry about this report, please raise from the date of receipt of the report within 30 days, overdue will not be accept. Items marked with "n" means they are not accredited by CNAS (if with CNAS logo), "s" means the item of subcontractor. All or Part of items of this report are not in the accredited scope of CMA and cannot be as domestic social impartiality proof data.

